

Date : 2021/08/10

Delta Electronics, Inc.
3, Tungyuan Road,
Chungli Industrial Zone,
Taoyuan City 32063
Taiwan, R.O.C.
Attn:

Re. : CU US + Canada Certificate

Type of Equipment : SWITCHING POWER SUPPLY
Model Designation : See Certificate
Certificate No. : CU 72210649 0003
File No. : CN216741 005
Engineer/Contact : Andreas Klinker
Standard(s) : CAN/CSA-C22.2 No. 60601-1:14
ANSI/AAMI ES60601-1:2005+A2 (R2012) +A1

Dear Madame or Sir,

The above referenced technical equipment has been tested and was found to be in compliance with the listed test requirement(s). Enclosed, please find the TUV Rheinland approval document No. CU 72210649 0003. It authorizes you to label the listed product(s) with the TUV Rheinland Mark identified in the approval document. For compliance, the Test Mark must be on the approved unit.

Your product is subject to regular factory follow-up inspections as well as annual certificate and factory registration fees.

In using the TUV Rheinland Mark you are obligated to comply with the TUV Rheinland of North America Service Agreement.

If we can be of any further assistance to you, please do not hesitate to contact us.

Sincerely yours,
Certification Body


Dipl.-Ing. W. Hsu
QA Certification Officer

Enclosure

Certificate



Certificate no.

CU 72210649 03

License Holder:

Delta Electronics, Inc.
3, Tungyuan Road,
Chungli Industrial Zone,
Taoyuan City 32063
Taiwan

Manufacturing Plant:

Delta Electronics (Jiangsu) Ltd.
215200, No.1688, Jiangxing East Road,
Wujiang Economic and Technological
Development Zone
Suzhou City, Jiangsu Province
P.R. China

Test report no.: USA-AK CN216741 005

Client Reference: Mr. Sam Tien

Tested to: CAN/CSA-C22.2 No. 60601-1:14
ANSI/AAMI ES60601-1:2005+A2 (R2012) +A1

Certified Product: SWITCHING POWER SUPPLY


License Fee - Units

Addition

Model Designation: MEU-650A12T X1X2X3
MEU-650Z12TA X1X2
(X1, X2, X3 = 0-9, A-Z, a-z, /, \, +, -, blank)

Appendix: 1, 1-25

Licensed Test mark:



Date of Issue
(day/mo/yr)
09/08/2021

Date : 2021/06/24

Delta Electronics, Inc.
3, Tungyuan Road,
Chungli Industrial Zone,
Taoyuan City 32063
Taiwan, R.O.C.
Attn:

Re. : CU US + Canada Certificate

Type of Equipment : SWITCHING POWER SUPPLY
Model Designation : See Certificate
Certificate No. : CU 72210649 0002
File No. : CN216741 004
Engineer/Contact : Andreas Klinker
Standard(s) : CAN/CSA-C22.2 No. 60601-1:14
ANSI/AAMI ES60601-1:2005+A2 (R2012) +A1

Dear Madame or Sir,

The above referenced technical equipment has been tested and was found to be in compliance with the listed test requirement(s). Enclosed, please find the TUV Rheinland approval document No. CU 72210649 0002. It authorizes you to label the listed product(s) with the TUV Rheinland Mark identified in the approval document. For compliance, the Test Mark must be on the approved unit.

Your product is subject to regular factory follow-up inspections as well as annual certificate and factory registration fees.

In using the TUV Rheinland Mark you are obligated to comply with the TUV Rheinland of North America Service Agreement.

If we can be of any further assistance to you, please do not hesitate to contact us.

Sincerely yours,
Certification Body



Dipl.-Ing. W. Hsu
QA Certification Officer

Enclosure

Certificate



Certificate no.

CU 72210649 02

License Holder:

Delta Electronics, Inc.
3, Tungyuan Road,
Chungli Industrial Zone,
Taoyuan City 32063
Taiwan

Manufacturing Plant:

Delta Electronics (Jiangsu) Ltd.
215200, No.1688, Jiangxing East Road,
Wujiang Economic and Technological
Development Zone
Suzhou City, Jiangsu Province
P.R. China

Test report no.: USA-AK CN216741 004

Client Reference: Sam Tien

Tested to:

CAN/CSA-C22.2 No. 60601-1:14
ANSI/AAMI ES60601-1:2005+A2 (R2012) +A1

Certified Product: SWITCHING POWER SUPPLY

License Fee - Units

Addition

Model Designation: MEU-650A24T X1X2X3
MEU-650A48T X1X2X3
MEU-650Z24TA X1X2
MEU-650Z48TA X1X2

(X1, X2, X3 = 0-9, A-Z, a-z, /, \, +, -, blank)


Change

Following Model Designation no longer valid:

MEU-650A24T

Appendix: 1, 1-18

Licensed Test mark:



Date of Issue

(day/mo/yr)

24/06/2021

Date : 23.02.2021

Delta Electronics, Inc.
3, Tungyuan Road,
Chungli Industrial Zone,
Taoyuan City 32063
Taiwan, R.O.C.
Attn:

Re. : CU US + Canada Certificate

Type of Equipment : SWITCHING POWER SUPPLY
Model Designation : See Certificate
Certificate No. : CU 72210649 0001
File No. : CN216741 001
Engineer/Contact : Andreas Klinker
Standard(s) : CAN/CSA-C22.2 No. 60601-1:14
ANSI/AAMI ES60601-1:2005+A2 (R2012) +A1

Dear Madame or Sir,

The above referenced technical equipment has been tested and was found to be in compliance with the listed test requirement(s). Enclosed, please find the TUV Rheinland approval document No. CU 72210649 0001. It authorizes you to label the listed product(s) with the TUV Rheinland Mark identified in the approval document. For compliance, the Test Mark must be on the approved unit.

Your product is subject to regular factory follow-up inspections as well as annual certificate and factory registration fees.

In using the TUV Rheinland Mark you are obligated to comply with the TUV Rheinland of North America Service Agreement.

If we can be of any further assistance to you, please do not hesitate to contact us.

Sincerely yours,
Certification Body


Dipl.-Ing. W. Hsu
QA Certification Officer

Enclosure

Certificate



Certificate no.

CU 72210649 01

License Holder:

Delta Electronics, Inc.
3, Tungyuan Road,
Chungli Industrial Zone,
Taoyuan City 32063
Taiwan

Manufacturing Plant:

Delta Electronics (Jiangsu) Ltd.
215200, No.1688, Jiangxing East Road,
Wujiang Economic and Technological
Development Zone
Suzhou City, Jiangsu Province
P.R. China

Test report no.: USA-AK CN216741 001

Client Reference: Sam Tien

Tested to:

CAN/CSA-C22.2 No. 60601-1:14
ANSI/AAMI ES60601-1:2005+A2 (R2012) +A1

Certified Product: SWITCHING POWER SUPPLY

License Fee - Units

Model Designation: MEU-650A24T

Rated Voltage: AC 100-120V, 50-60Hz
AC 200-240V, 50-60Hz

Rated Current: 8A

Output Ratings: see construction dataform

max. total power: see construction dataform

Protection Class: Class I

Rated Ambient Temperature: 50°C max.(full load)
70°C max.(de-rating half load)

Appendix: 1, 1-17

Licensed Test mark:



Date of Issue

(day/mo/yr)

23/02/2021